



PUBLIC HEALTH ENGINEERING DEPARTMENT
STATE WATER TESTING LABORATORY, KARNAL

Website :- <https://phedharyana.gov.in/>

Date:- 23/04/2025

Memo No:- 2911

Subject : **Bacteriological Examination Report of Water Sample**

Sample Details		Date & Time
Sample ID:	KNL/01304/4/2025	Collected By
Sample Classification:	Private	ANIL KUMAR KUN.....9896628688
Sender:	PM SHRI GSSS BALRAJPUTAN	Collection Date
Location :-	, VPO BALRAJPUTAN	21/04/2025 at 10:21
Latitude :-	76.810978	Received at lab
Longitude :-	29.549422	22/04/2025 at 16:12
		Tested at Lab
		23/04/2025

Test Result (As Per BIS 10500 :2012)					Protocol Used
S.No.	Tested Parameter	Result	Requirement (Acceptable Limit)		
1	Total Coliform per 100 ml	0.0	Shall not be detectable in any 100 ml sample		IS 15185:2016/APHA 9223B-2023/APHA 9221C-2023

Remarks :-

- * The Results Given above are related to the sample as received and tested in KARNAL lab.
- * The test report can't be regenerated in whole or part there of without written permission of Competent Authority.
- * The test report can't be used for any publicity or any legal purpose.
- * Reliability of water sample sample lies with sender/collector of water sample.
- * To check the sample report online Scan the QR Code below.

Water Sample has been found **"CONFORMING"** to Limits set in BIS10500:2012 for the tested parameters.
Based on the tested parameters, the water sample is found **"POTABLE"**



(Signature)
AMIT KUMAR, CHIEF CHEMIST 23/04/25
PUBLIC HEALTH ENGINEERING DEPARTMENT
STATE WATER TESTING LABORATORY, KARNAL
email: chiefchemisthr@phedharyana.gov.in

-----End of Report-----

(Signature)